

6390 Joyce Drive
100
Golden, CO 80403

Phone 303-940-0033
Fax 866-283-0269
www.phenova.com

July 5, 2018

Ms. Nichole Jones
BoJhun Environmental Services
222 S. Main Street
Fostoria, OH 44830
United States

Dear Nichole,

Thank you for participating in the WPM0618 Water Pollution Proficiency Testing Study. Enclosed is your final report, which has been carefully reviewed by the PT specialists at Phenova. This report presents some additional sections and features which will give you more information on the study data for the standards that were reported by your laboratory and more information regarding the overall performance of your laboratory in relation to other study participants.

The enclosed report is in two sections, the "Evaluation Report" and the "Statistical Report". The "Evaluation Report" contains the evaluation of your reported data as well the reported method, analysis date and analyst. The "Evaluation Report" is an updated and enhanced version of the reports you previously received at the conclusion of our PT studies. The "Statistical Report" contains your passing percentage as well as statistics from the study for the analytes reported by your laboratory.

For any analyte falling outside the established acceptance limits, our PT management staff would like to assist you in determining the most appropriate course of corrective action for your facility. Please contact us at any time if we may be of service to you.

Thank you again for participating in the WPM0618 Water Pollution Proficiency Testing Study. We appreciate working with you and look forward to our next study. If you have any questions, please call us at 866-942-2978.



Report Definitions:

Assigned Value

The Assigned Value is determined from the study mean, gravimetric and volumetric true concentration of an analyte to be analyzed, calculation and/or an appropriate reference value as stipulated in the EPA National Standards for Water Proficiency Testing Studies Criteria Document (current version), the National Environmental Laboratory Accreditation Conference (NELAC) criteria (ref: NELAC FOT tables, NELAC PT Committee) and other documents distributed by accrediting

Evaluation Limits

Acceptance Limits are derived from fixed limits, coefficients, constants and calculations stipulated in the EPA National Standards for Water Proficiency Testing Studies Criteria Documents (current version), the National Environmental Laboratory Accreditation Conference (NELAC) criteria (ref: NELAC PT FOT tables, NELAC PT Committee) and other documents distributed by accrediting agencies as applicable.

Evaluation

- Acceptable* The reported value falls within the Acceptance Limits.
- Not Acceptable* The reported value falls outside the Acceptance Limits.
- No Evaluation* The reported value is non-numeric and can not be evaluated.
- NR* As required by the NELAC standards and requested by state authorities, any analyte purchased but not reported by your facility is listed as NR (Not Reported).

Study Mean and Standard Deviation

The mean and standard deviation are calculated from the study data using robust statistical calculations when possible. Other statistical calculations may be used if robust statistical calculations are not possible. The displayed values are independent of any statistical calculations required for the calculation of the analyte evaluation limits.

This report must not be reproduced except in full, without the written approval of Phenova.

This report must not be used to claim product certification, approval, or endorsement by the American Association for Laboratory Accreditation (A2LA) or any agency of the federal government.



6390 Joyce Drive
100
Golden, CO 80403

Phone 303-940-0033
Fax 866-283-0269
www.phenova.com

State Regulatory Agency Request(s):

Phenova has been authorized to send a copy of your WPM0618 final results to the following state agencies:

OH

Other Regulatory Agency Request(s):

No requests were made to send a copy of your WPM0618 final results to an agency.

Third Party Request(s):

No requests were made to send a copy of your WPM0618 final results to a third party.

6390 Joyce Drive Phone 303-940-0033
100 Fax 866-283-0269
Golden, CO 80403 www.phenova.com

Final Report - Water Pollution Proficiency Testing

Study: WPM0618

Opening Date: June 12, 2018 - Closing Date: June 28, 2018

Laboratory: BoJhun Environmental Services
222 S. Main Street
Fostoria, OH 44830
UNITED STATES

Contact: Ms. Nichole Jones, Owner
419-436-7066

EPA Lab ID: OH01333

Micro (PT-MIC-WP) Lot #: 5281-12

NELAC Code	Analyte	Analysis Date	Analyst	Method Code	Method Description	Units	Assigned Value	Result	Acceptance Limits	Evaluation
2530	Fecal Coliform, MF					CFU/100 mL	659		332 - 1310	NR
2525	E. Coli, MF					CFU/100 mL	720		317 - 1640	NR
2500	Total Coliform, MF					CFU/100 mL	880		295 - 2630	NR
2530	Fecal Coliform, MPN					MPN/100 mL	828		378 - 1820	NR
2525	E. Coli, MPN	6/13/2018	NJ/AS	20211614	SM 9223 B (Colilert Quanti-Tray)-2004 2004	MPN/100 mL	940	1299.7	431 - 2050	Acceptable
2500	Total Coliform, MPN	6/13/2018	NJ/AS	20211614	SM 9223 B (Colilert Quanti-Tray)-2004 2004	MPN/100 mL	882	1299.7	479 - 1630	Acceptable



6390 Joyce Drive
 # 100
 Golden, CO 80403

Phone 303-940-0033
 Fax 866-283-0269
 www.phenova.com

PT Study Score Report

EPA ID: OH01333
 Laboratory: BoJhun Environmental Services
 222 S. Main Street
 Fostoria, OH 44830
 United States

Study Number	Number of Reported Results	Number of Passing Results	Percent Passing
WPM0618	2	2	100%

Report Definitions:

- Number of Reported Results** The number of results reported which could be evaluated. Results receiving an evaluation of "NR" or "No Evaluation" are not included.
- Number of Passing Results** The number of results reported receiving an evaluation of "Acceptable" or "Check for Error".
- Percent Passing** The percentage of results reported receiving an evaluation of "Acceptable" or "Check for Error".



6390 Joyce Drive
100
Golden, CO 80403

Phone 303-940-0033
Fax 866-283-0269
www.phenova.com

Statistical Summary Report - Water Pollution Proficiency Testing

Opening Date: June 12, 2018 - Closing Date: June 28, 2018

Laboratory: BoJhun Environmental Services
222 S. Main Street
Fostoria, OH 44830
UNITED STATES

Contact: Ms. Nichole Jones, Owner
419-436-7066

EPA Lab ID: OH01333

Micro (PT-MIC-WP)						Lot #: 5281-12
NELAC Code	Analyte	Units	Study Mean	Study Standard Deviation	Low Acceptance Limit %	High Acceptance Limit %
2530	Fecal Coliform, MF	CFU/100 mL	675	161	50.4	199
2525	E. Coli, MF	CFU/100 mL	688	196	44.0	228
2500	Total Coliform, MF	CFU/100 mL	897	228	33.5	299
2530	Fecal Coliform, MPN	MPN/100 mL	853	205	45.7	220
2525	E. Coli, MPN	MPN/100 mL	970	251	45.9	218
2500	Total Coliform, MPN	MPN/100 mL	899	178	54.3	185